

# Elecsys CA 72-4

cobas®

REF	$\Sigma$	SYSTEM
11776258 122	100	MODULAR ANALYTICS E170 cobas e 411 cobas e 601 cobas e 602

## English

### System information

For **cobas e 411** analyzer: test number 360  
For MODULAR ANALYTICS E170, **cobas e 601** and **cobas e 602** analyzers: Application Code Number 055

### Please note

The measured CA 72-4 value of a patient's sample can vary depending on the testing procedure used. The laboratory finding must therefore always contain a statement on the CA 72-4 assay method used. CA 72-4 values determined on patient samples by different testing procedures cannot be directly compared with one another and could be the cause of erroneous medical interpretations. If there is a change in the CA 72-4 assay procedure used while monitoring therapy, then the CA 72-4 values obtained upon changing over to the new procedure must be confirmed by parallel measurements with both methods.

### Intended use

Immunoassay for the in vitro quantitative determination of CA 72-4 in human serum and plasma. The assay in particular serves as an aid in the therapeutic monitoring of carcinomas of the stomach and ovaries.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

### Summary

The tumor associated glycoprotein (TAG) 72, also known as CA 72-4 is a mucin protein of high molecular weight (approximately 200-400 kD) and found on the surface of many cancer cells, including stomach, ovary, breast, colon and pancreatic cells.<sup>1</sup> An antibody construct directed against TAG 72 has been proposed as an anti-tumor agent against ovarian and prostate cancer.<sup>2</sup>

Elevated serum levels are primarily found in gastric cancer patients,<sup>3,4</sup> but can also be found in certain non-malignant diseases like pneumonia, pancreatitis, liver cirrhosis and ovarian cysts.<sup>5</sup> The most important advantage of CA 72-4 is its ability to discriminate between malignant and non-malignant diseases.<sup>3,6</sup>

#### Gastric cancer:

For gastric cancer, a diagnostic sensitivity of 33 % was reported for CA 72-4, which increased to 66 % when a combination of CA 72-4, CEA, CA-125 and CA 19-9 was used.<sup>7</sup> There is a correlation between the stage of disease and the level of CA 72-4.<sup>8</sup> Monitoring treatment and disease course in patients with gastric cancer is the main indication for CA 72-4 as a first-line biomarker or in conjunction with a second marker (CEA or CA 19-9).<sup>9</sup> It was shown that preoperative serum levels of CA 72-4 had the best predictive value in indicating advanced disease in patients diagnosed with gastric cancer.<sup>10</sup> After surgical intervention, CA 72-4 levels return to normal and remain within the normal range in cases where tumor tissue is no longer present.<sup>11</sup> In 70 % of relapse cases, CA 72-4 increases prior or concurrently with clinical diagnosis of the relapse.<sup>5</sup>

#### Ovarian cancer:

A diagnostic sensitivity of 47-80 % has been reported in ovarian carcinoma.<sup>12</sup> Especially for mucinous ovarian cancer, the diagnostic sensitivity of CA 72-4 is greater than that of CA 125. Combined use of the two markers provides an additive diagnostic sensitivity of 73 % for primary diagnosis (CA 125 alone: 60 %) and 67 % for monitoring purposes (CA 125 alone: 60 %).<sup>8</sup>

#### Colorectal cancer:

Diagnostic sensitivity for colorectal carcinoma is 20-41 %.<sup>13</sup> There is a correlation with the clinical staging by Dukes.<sup>14</sup> Diagnostic specificity of CA 72-4 to benign diseases of the colon is 98 %.<sup>15</sup> After complete resection a marked drop in CA 72-4 occurs. In long-term controls, the CA 72-4 concentration remains elevated when a residual tumor is present.

Combined use of CA 72-4 and CEA increases diagnostic sensitivity from 78 % to 87 % in post-operative relapse controls.<sup>8</sup>

The Elecsys CA 72-4 assay utilizes the following two monoclonal antibodies to detect the mucin, TAG 72:<sup>16</sup>

- B72.3 monoclonal antibody, which has been raised against a membrane-enriched extract of mammary carcinoma metastases<sup>17</sup> and
- CC49 monoclonal antibody, specific to highly-purified TAG 72.

### Test principle

Sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: 30 µL of sample, a biotinylated monoclonal CA 72-4-specific antibody (CC49), and a monoclonal CA 72-4-specific antibody (B72.3) labeled with a ruthenium complex<sup>a)</sup> react to form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode or e-barcode.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)<sub>3</sub><sup>2+</sup>)

### Reagents - working solutions

The reagent rackpack is labeled as CA72-4.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:  
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-CA 72-4-Ab~biotin (gray cap), 1 bottle, 8 mL:  
Biotinylated monoclonal anti-CA 72-4 antibody (CC49; mouse) 1 mg/L; phosphate buffer 100 mmol/L, pH 6.8; preservative.
- R2 Anti-CA 72-4-Ab~Ru(bpy)<sub>3</sub><sup>2+</sup> (black cap), 1 bottle, 8 mL:  
Monoclonal anti-CA 72-4 antibody (B72.3; mouse) labeled with ruthenium complex 6 mg/L; phosphate buffer 100 mmol/L, pH 6.8; preservative.

### Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

### Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

### Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

# Elecsys CA 72-4

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks
on the analyzers	8 weeks

## Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K<sub>2</sub>-EDTA and K<sub>3</sub>-EDTA plasma.

Criterion: Recovery within 90-110 % of serum value or slope 0.9-1.1 + intercept within  $\pm 2 \times$  analytical sensitivity (LDL) + coefficient of correlation  $> 0.95$ .

Stable for 24 hours at 20-25 °C, 30 days at 2-8 °C, 90 days at -20 °C ( $\pm 5$  °C). Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

## Materials provided

See "Reagents – working solutions" section for reagents.

## Materials required (but not provided)

- [REF] 11776274122, CA 72-4 CalSet, for 4 x 1 mL
- [REF] 11776452122, PreciControl Tumor Marker, for 4 x 3 mL
- [REF] 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF] 03183971122, Diluent Universal, 2 x 36 mL sample diluent
- General laboratory equipment
- MODULAR ANALYTICS E170 or **cobas e** analyzer

Accessories for **cobas e** 411 analyzer:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean
- [REF] 11706802001, AssayCup, 60 x 60 reaction cups
- [REF] 11706799001, AssayTip, 30 x 120 pipette tips
- [REF] 11800507001, Clean-Liner

Accessories for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 12102137001, AssayTip/AssayCup, 48 magazines x 84 reaction cups or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Accessories for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

## Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers (except for the **cobas e** 602 analyzer).

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

## Calibration

Traceability: This method has been standardized against the Enzymun-Test CA 72-4 method.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

**Calibration frequency:** Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 12 weeks when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

## Quality control

For quality control, use PreciControl Tumor Marker.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

## Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in U/mL or kU/L).

## Limitations - interference

The assay is unaffected by icterus (bilirubin  $< 1129$   $\mu\text{mol/L}$  or  $< 66$  mg/dL), hemolysis (Hb  $< 1.4$  mmol/L or  $< 2.2$  g/dL), lipemia (Intralipid  $< 1500$  mg/dL) and biotin ( $< 246$  nmol/L or  $< 60$  ng/mL).

Criterion: Recovery within  $\pm 10$  % of initial value.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e.  $> 5$  mg/day) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 1500 IU/mL.

There is no high-dose hook effect at CA 72-4 concentrations up to 15000 U/mL.

In vitro tests were performed on 27 commonly used pharmaceuticals. No interference with the assay was found.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

# Elecsys CA 72-4



For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

## Limits and ranges

### Measuring range

0.200-300 U/mL (defined by the lower detection limit and the maximum of the master curve). Values below the lower detection limit are reported as < 0.200 U/mL. Values above the measuring range are reported as > 300 U/mL (or up to 600 U/mL for 2-fold diluted samples).

### Lower limits of measurement

#### Lower detection limit of the test

Lower detection limit: < 0.20 U/mL

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying two standard deviations above that of the lowest standard (master calibrator, standard 1 + 2 SD, repeatability study, n = 21).

### Dilution

Samples with CA 72-4 concentrations above the measuring range can be diluted with Diluent Universal. The recommended dilution is 1:2 (either automatically by the analyzers or manually). The concentration of the diluted sample must be > 150 U/mL.

After manual dilution, multiply the result by the dilution factor.

After dilution by the analyzers, the software automatically takes the dilution into account when calculating the sample concentration.

**Please note:** A sample dependent non-linearity of dilutions has been observed for samples which are outside the measuring range.

### Expected values

Extended studies with the Elecsys CA 72-4 assay in clinical centers in Belgium, Germany, and Roche-internal studies gave the following results for a total of 635 healthy individuals:

6.9 U/mL (95 % percentile)

5.6-8.2 U/mL (95 % confidence range of the percentile)<sup>18</sup>

Status: Elecsys CA 72-4 multicenter evaluation; study No. B99P026, 7/2001

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

### Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

### Precision

Precision was determined using Elecsys reagents, pooled human sera and controls in accordance with a modified protocol (EP5-A) of the CLSI (Clinical and Laboratory Standards Institute): 6 times daily for 10 days (n = 60); repeatability on MODULAR ANALYTICS E170 analyzer, n = 21. The following results were obtained:

cobas e 411 analyzer						
	Repeatability			Intermediate precision		
Sample	Mean U/mL	SD U/mL	CV %	Mean U/mL	SD U/mL	CV %
Human serum 1	3.12	0.06	2.0	3.39	0.12	3.6
Human serum 2	19.2	0.34	1.8	20.1	0.85	4.2
Human serum 3	107	2.26	2.1	119	5.79	4.9
PreciControl TM <sup>b</sup> 1	2.85	0.06	2.1	2.87	0.08	2.9
PreciControl TM2	40.4	0.98	2.4	43.0	2.12	4.9

b) TM = Tumor Marker

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers						
	Repeatability			Intermediate precision		
Sample	Mean U/mL	SD U/mL	CV %	Mean U/mL	SD U/mL	CV %
Human serum 1	3.23	0.04	1.4	3.25	0.07	2.2

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers						
	Repeatability			Intermediate precision		
Sample	Mean U/mL	SD U/mL	CV %	Mean U/mL	SD U/mL	CV %
Human serum 2	18.8	0.50	2.8	137	4.90	3.6
Human serum 3	149	3.93	2.8	18.6	0.46	2.5
PreciControl TM1	3.90	0.05	1.0	3.78	0.12	3.1
PreciControl TM2	33.3	0.38	1.0	32.5	0.98	3.0

### Method comparison

A comparison of the Elecsys CA 72-4 assay (y) with the Enzymun-Test CA 72-4 method (x) using clinical samples gave the following correlations:

Number of samples measured: 144

Passing/Bablok<sup>19</sup> Linear regression

y = 0.93x - 1.59 y = 0.95x - 1.43

r = 0.877 r = 0.954

The sample concentrations were between approximately 0.3 and approximately 87 U/mL.

### Analytical specificity

The Elecsys CA 72-4 tumor marker assay is based on the monoclonal B72.3 and CC49 antibodies which are only available from Fujirebio Diagnostics, its licensees and its representatives. The performance characteristics of testing procedures using these antibodies cannot be assumed for test methods using other antibodies.

### Functional sensitivity

1.0 U/mL

The functional sensitivity is the lowest analyte concentration that can be reproducibly measured with an intermediate precision CV of < 20 %.

### References

- Muraro R, Kuroki M, Wunderlich D, et al. Generation and characterization of B72.3 2nd generation monoclonal antibodies reactive with the tumor-associated glycoprotein 72 antigen. *Cancer Res* 1988;48(16):4588-4596.
- Scott MA, Akhurst T, Lee F-T, et al. Phase I safety and biodistribution study of 124I-PEG-AVP0458 diabody in patients with TAG-72 positive ovarian and prostate cancer. In: Proceedings of the 106th Annual Meeting of the American Association for Cancer Research; 2015 Apr 18-22; Philadelphia, PA. Philadelphia (PA): AACR; *Cancer Res* 2015;75(15):Abstract nr CT238.
- Guadagni F, Roselli M, Cosimelli M, et al. CA 72-4 Serum Marker - A New Tool in the Management of Carcinoma Patients. *Cancer Invest* 1995;13(2):227-238.
- Filella X, Fuster J, Molina R, et al. TAG-72, CA 19-9 and CEA as tumor markers in gastric cancer. *Acta Oncologica* 1994;33(7):747-751.
- Filella X, Molina R, Jo J, et al. Tumor associated glycoprotein 72 (TAG 72) levels in patients with non-malignant and malignant disease. *Bull Cancer* 1992;79:271-277.
- Heptner G, Domschke S, Domschke W. Comparison of CA 72-4 with CA 19-9 and Carcinoembryonic Antigen in the Serodiagnosis of Gastrointestinal Malignancies. *Scand J Gastroenterol* 1989;24:745-750.
- Yang A-P, Liu J, Lei H-Y, et al. CA 72-4 combined with CEA, CA-125 and CA 19-9 improves the sensitivity for the early diagnosis of gastric cancer. *Clin Chim Acta* 2014;437:183-86.
- Lamerz R. CA 72-4 (TAG-72). In: Thomas L (ed.). *Clinical Laboratory Diagnosis*, TH-Books, Frankfurt, 1st English Edition 1998:952-955.
- Shimada H, Nole T, Ohashi M, et al. Clinical significance of serum tumor markers for gastric cancer: a systematic review of literature by the Task Force of the Japanese Gastric Cancer Association. *Gastric Cancer* 2014;17:26-33.
- Cidon EU, Bustamante R. Gastric cancer: tumor markers as predictive factors for preoperative staging. *J. Gastrointest Cancer* 2011; 42: 127-130.

# Elecsys CA 72-4



- 11 Reiter W, Stieber P, Reuter C, et al. Prognostic Value of Preoperative Serum Levels of CEA, CA 19-9 and 72-4 in Gastric Carcinoma. *Anticancer Res* 1997;17:2903-2907.
- 12 Hasholzner U, Baumgartner L, Stieber P, et al. Clinical significance of the tumor markers CA 125 II and CA 72-4 in ovarian carcinoma. *Int J Cancer (Pred Oncol)* 1996;69(4):329-334.
- 13 Guadagni F, Roselli M, Cosimelli M, et al. TAG-72 Expression and its Role in the Biological Evaluation of Human Colorectal Cancer. *Anticancer Res* 1996;16:2141-2148.
- 14 Sila A, Roselli M, Cosimelli M, et al. Clinical Utility of CA 72-4 Serum Marker in the Staging and Immediate Post-surgical Management of Gastric Cancer Patients. *Anticancer Res* 1996;16:2241-2248.
- 15 Stieber P, Fateh-Moghadam A, Wädlich H, et al. CA 72-4: A new tumour marker for stomach cancer. In Klapdor R, ed. *Recent results in tumor diagnosis and therapy*. München: Zuckschwerdt 1990:23-26.
- 16 Johnson VG, Schlom J, Paterson AJ, et al. Analysis of a human tumor-associated glycoprotein (TAG-72) identified by monoclonal antibody B72.3. *Cancer Res* 1986;46:850-857.
- 17 Colcher D, Horan Hand P, Nuti M, et al. A Spectrum of monoclonal antibodies reactive with human mammary tumor cells. *Proc Natl Acad Sci* 1981;78(5):3199-3208.
- 18 Hahn GJ, Meeker WQ. *Statistical Intervals: A Guide for Practitioners*. John Wiley & Sons, Inc. New York 1991.
- 19 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.



**FUJIREBIO**  
Diagnostics, Inc.

CA 72-4 is a registered trademark of  
Fujirebio Diagnostics, Inc.

## Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see <https://usdiagnostics.roche.com> for definition of symbols used):

	CONTENT	Contents of kit
	SYSTEM	Analyzers/Instruments on which reagents can be used
	REAGENT	Reagent
	CALIBRATOR	Calibrator
		Volume after reconstitution or mixing
	GTIN	Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2017, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)

