

REF 04528417 190

→ 3 x 2 mL Calibrator

English

System information

For use on Roche/Hitachi MODULAR and **cobas c** analyzers the calibrator code is 674.

For use on the Roche/Hitachi **cobas c** 513 analyzer the calibrator code is 30674.

For use on COBAS INTEGRA analyzers the system ID is 07 6852 9.

Intended use

C.f.a.s. (Calibrator for automated systems) HbA1c is for use in the calibration of quantitative Roche methods on Roche clinical chemistry analyzers as specified in the enclosed value sheets.

Summary

C.f.a.s. HbA1c is a lyophilized calibrator based on hemolyzed sheep blood. The concentrations of the calibrator components have been adjusted to ensure optimal calibration of the appropriate Roche methods on clinical chemistry analyzers.

Some methods specified in the relevant value sheet may not be available in all countries.

Reagents - working solutions

Reactive components:

Hemolyzed sheep blood with chemical additives and material of biological origin as specified. The origin of the biological additives is as follows:

Analyte	Origin
Hemoglobin	Sheep blood
HbA1c	Human blood

Non-reactive components:

Preservatives and stabilizers

The concentrations of the calibrator components are lot-specific. The exact calibrator values are given in the electronically available or enclosed value sheets.

C.f.a.s. HbA1c calibrator values are reagent lot matched (except for Roche/Hitachi MODULAR, Roche/Hitachi 902 and **cobas c** 111 analyzers). For each application and each combination of C.f.a.s. HbA1c calibrator lot and Tina-quant Hemoglobin A1c Gen.2 (for COBAS INTEGRA analyzers), Tina-quant Hemoglobin A1c Gen.3 (for **cobas c** 311/501 and **cobas c** 502 analyzers) or Tina-quant Hemoglobin A1cDx Gen.3 (for the **cobas c** 513 analyzer) reagent lot the exact calibrator values are given in the electronically available or enclosed value sheet.

The values are also encoded in the enclosed calibrator barcode sheets for Roche/Hitachi MODULAR, COBAS INTEGRA and **cobas c** 111 analyzers.

For the **cobas c** analyzers (except for the **cobas c** 111 analyzer) the values are encoded in electronic files sent via the **cobas** link to the analyzers.

Calibrator values

The calibrator values were determined using the method stated in the electronically available or enclosed value sheets. Determinations were performed under strictly standardized conditions on Roche analyzers using Roche system reagents and the Roche master calibrator.

The calibrator values were obtained via single determinations performed in different laboratories, in several separate runs. The calibrator value specified is the mean of all values obtained.

Traceability information is given in the relevant Method Sheets for the system reagent.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

For USA: For prescription use only.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

2-Methyl-2H-isothiazol-3-one hydrochloride, 26172-54-3

EUH 208 May produce an allergic reaction.



Danger

H314 Causes severe skin burns and eye damage.

Prevention:

P260 Do not breathe dust or mist.

P280 Wear protective gloves/ protective clothing/ eye protection/ face protection.

Response:

P303 + P361 + P353 IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower.

P304 + P340 + P310 IF INHALED: Remove person to fresh air and keep comfortable for breathing. Immediately call a POISON CENTER or doctor/physician.

P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

Product safety labeling primarily follows EU GHS guidance.

Contact phone: all countries: +49-621-7590, USA: 1-800-428-2336

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A. However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{1,2}

Handling

Carefully open one bottle, avoiding the loss of lyophilizate, and pipette in exactly 2.0 mL of distilled/deionized water. Carefully close the bottle and dissolve the contents completely by occasional gentle swirling within 30 minutes. Avoid the formation of foam.

The enclosed barcoded labels are intended exclusively for Roche/Hitachi MODULAR automated analyzers and **cobas c** systems (except for the **cobas c** 513 analyzer) to identify the calibrator. Attach the barcoded labels to the tubes carrying the sample cups containing the calibrator material.

Storage and stability

Store at 2-8 °C.

Criterion for the stability data stated by Roche:

Recovery within ± 10 % of initial value.

Stability of the lyophilized calibrator at 2-8 °C:

up to the stated expiration date.

Stability of the components in the reconstituted calibrator:

at 15-25 °C 8 hours

at 2-8 °C 2 days

at (-15)-(-25) °C 3 months (when frozen once)

Store calibrator tightly capped when not in use.

Materials provided

- See "Reagents – working solutions" section

C.f.a.s. HbA1c



- Barcoded labels

Materials required (but not provided)

- Roche system reagents and clinical chemistry analyzers
- General laboratory equipment

Assay

Use C.f.a.s. HbA1c as specified in the relevant Method Sheet for the system reagents.

References

- 1 Occupational Safety and Health Standards: Bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- 2 Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Calibrator
	Volume after reconstitution or mixing
	Global Trade Item Number

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C, COBAS INTEGRA and TINA-QUANT are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

Distribution in USA by:

Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336

